Patient Instruction Sheet

∣ trackıt mk3[¬]

Upon arrival at home, after having the equipment and electrodes fitted to you, please observe the following instructions to help ensure a successful recording:

DO:



Keep the equipment in its bag to protect it from ingress of water.



- Ensure that the bag and straps are worn over clothing to prevent any possibility of skin irritation.
- Take care to avoid entanglement of cables or straps. Entangled cables or straps could result in personal injury or strangulation.
- Press the event button (one single quick press) if you feel anything unusual
 - Contact the number, provided above, if the EEG electrodes become loose or detached, or if you experience any problems with the equipment.

Environmental operating conditions

Temperature	+5°C - +40°C (+41°F to +104°F)
Humidity	15% - 90%
Protection class	IP22, when contained in the bag.
	<i>IP22 Explanation</i> 1st digit (2): Protected against foreign objects greater than 12.5mm. 2nd digit (2): Protected against dripping water when the equipment is tilted up to 15°.

For help, call:


DON'T:



X

The equipment requires no patient intervention during the recording. Once the recording has been set up, do not touch / tamper with the laptop during a recording.

Do not remove the batteries from the amplifier.



- Do not use mobile phones in close proximity to the equipment.
- Do not chew gum or other sweets as this produces artefact signals.
- Do not bath or shower.
- Do not subject the equipment to heat or moisture. See Environmental operating conditions.
- Do not allow pets or children to interfere with the equipment or the cables.



CAUTION: Federal (USA) law restricts this device to sale by or on the order of a physician.

51329-006 - Mk3 EEG Amplifier Patient Instruction Sheet Iss 1.1

Lifelines Ltd, 1 Tannery House. Tannery Lane, Woking, Surrey, GU23 7EF. UK Tel: +44 (0) 1483 224 245 · www.lifelinesneuro.com · sales@lifelinesneuro.com



